



Anova Laboratory, LLC
 1685 Tucker Road
 Hood River, OR 97031
 541.387.4020 (office)
 866.293.1337 (fax)

Preliminary Analytical Results

Sofresh Farms
 27135 S. Gribble Rd.
 Canby, OR 97013
 Michelle Haworth
 Lic Num: 020-100010499D8

Strain: Betty Hayes
 Matrix: Flower
 CTS Batch #: 1A4010300001E17000005929
 Lot Number: Betty Hayes 8/3/18
 Lot Size (lb): 7.65
 Client CTS Sample ID: SOF_6387
 Collected: 3/1/2019 13:11
 Received: 3/1/2019 16:32
 Anova Sample ID: 079190600201

Analysis Performed	Date Analyzed	Result	fn	Reporting Units	Reporting Criterion	Pass/Fail
Giese 2015						
Cannabidiol (CBD) by HPLC	3/7/2019	< 0.02		% Mass	n/a	n/a
Cannabidiolic acid (CBDA) by HPLC	3/7/2019	0.09		% Mass	n/a	n/a
Cannabinol (CBN) by HPLC	3/7/2019	0.26		% Mass	n/a	n/a
delta-9-Tetrahydrocannabinol (d9-THC) by HPLC	3/7/2019	6.08		% Mass	n/a	n/a
delta-9-Tetrahydrocannabinolic acid (d9-THCA) by HPLC	3/7/2019	27.0		% Mass	n/a	n/a
Total CBD (Calculated)	3/7/2019	0.08		% Mass	n/a	n/a
Total THC (Calculated)	3/7/2019	29.8		% Mass	n/a	n/a
AOAC Method 934.01						
Moisture (Loss on Drying)	3/5/2019	12.82		% Mass	15 max	Pass
AOAC Method 978.18						
Water Activity	3/5/2019	0.35		aW	0.65 max	Pass
AOAC 2007.01 and EN 15662: LC-MS/MS						
Abamectin (CAS# 71751-41-2)	3/7/2019	< 0.10		ppm (ug/g)	0.5 max	Pass
Acephate (CAS# 30560-19-1)	3/7/2019	< 0.05		ppm (ug/g)	0.4 max	Pass
Acequinocyl (CAS# 57960-19-7)	3/7/2019	< 0.20		ppm (ug/g)	2 max	Pass
Acetaminiprid (CAS# 135410-20-7)	3/7/2019	< 0.05		ppm (ug/g)	0.2 max	Pass
Aldicarb (CAS# 116-06-3)	3/7/2019	< 0.05		ppm (ug/g)	0.4 max	Pass
Azoxystrobin (CAS# 131860-33-8)	3/7/2019	< 0.05		ppm (ug/g)	0.2 max	Pass
Bifenazate (CAS# 149877-41-8)	3/7/2019	< 0.05		ppm (ug/g)	0.2 max	Pass
Boscalid (CAS# 188425-85-6)	3/7/2019	< 0.05		ppm (ug/g)	0.4 max	Pass
Carbaryl (CAS# 63-25-2)	3/7/2019	< 0.05		ppm (ug/g)	0.2 max	Pass
Carbofuran (CAS# 1563-66-2)	3/7/2019	< 0.05		ppm (ug/g)	0.2 max	Pass
Chlorantraniliprole (CAS# 500008-45-7)	3/7/2019	< 0.05		ppm (ug/g)	0.2 max	Pass

This report shall not be reproduced without the expressed written consent of ANOVA Laboratory, LLC.

These results pertain only to the representative sample submitted for analysis.

Anova Laboratory, LLC - OLCC Lic. No. 10010169ADI ORELAP Accreditation No. 4098

Client-Based Solutions and High-Quality, Rapid Results



Anova Laboratory, LLC
 1685 Tucker Road
 Hood River, OR 97031
 541.387.4020 (office)
 866.293.1337 (fax)

Preliminary Analytical Results

Sofresh Farms
 27135 S. Gribble Rd.
 Canby, OR 97013
 Michelle Haworth
 Lic Num: 020-100010499D8

Strain: Betty Hayes
 Matrix: Flower
 CTS Batch #: 1A4010300001E17000005929
 Lot Number: Betty Hayes 8/3/18
 Lot Size (lb): 7.65
 Client CTS Sample ID: SOF_6387
 Collected: 3/1/2019 13:11
 Received: 3/1/2019 16:32
 Anova Sample ID: 079190600201

Analysis Performed	Date Analyzed	Result	fn	Reporting Units	Reporting Criterion	Pass/Fail
Chlorpyrifos (CAS# 2921-88-2)	3/7/2019	< 0.05		ppm (ug/g)	0.2 max	Pass
Clofentezine (CAS# 74115-24-5)	3/7/2019	< 0.05		ppm (ug/g)	0.2 max	Pass
Cyfluthrin (CAS# 68359-37-5)	3/7/2019	< 0.20		ppm (ug/g)	1 max	Pass
Cypermethrin (CAS# 52315-07-8)	3/7/2019	< 0.10		ppm (ug/g)	1 max	Pass
Daminozide (CAS# 1596-84-5)	3/7/2019	< 0.10		ppm (ug/g)	1 max	Pass
DDVP (Dichlorvos) (CAS# 62-73-7)	3/7/2019	< 0.05		ppm (ug/g)	1 max	Pass
Diazinon (CAS# 333-41-5)	3/7/2019	< 0.05		ppm (ug/g)	0.2 max	Pass
Dimethoate (CAS# 60-51-5)	3/7/2019	< 0.05		ppm (ug/g)	0.2 max	Pass
Ethoprophos (CAS# 13194-48-4)	3/7/2019	< 0.05		ppm (ug/g)	0.2 max	Pass
Etofenprox (CAS# 80844-07-1)	3/7/2019	< 0.05		ppm (ug/g)	0.4 max	Pass
Etoxazole (CAS# 153233-91-1)	3/7/2019	< 0.05		ppm (ug/g)	0.2 max	Pass
Fenoxycarb (CAS# 72490-01-8)	3/7/2019	< 0.05		ppm (ug/g)	0.2 max	Pass
Fenpyroximate (CAS# 134098-61-6)	3/7/2019	< 0.05		ppm (ug/g)	0.4 max	Pass
Flonicamid (CAS# 158062-67-0)	3/7/2019	< 0.05		ppm (ug/g)	1 max	Pass
Fludioxonil (CAS# 131341-86-1)	3/7/2019	< 0.05		ppm (ug/g)	0.4 max	Pass
Hexythiazox (CAS# 78587-05-0)	3/7/2019	< 0.05		ppm (ug/g)	1 max	Pass
Imazalil (CAS# 35554-44-0)	3/7/2019	< 0.05		ppm (ug/g)	0.2 max	Pass
Imidacloprid (CAS# 138261-41-3)	3/7/2019	< 0.05		ppm (ug/g)	0.4 max	Pass
Kresoxim-methyl (CAS# 143390-89-0)	3/7/2019	< 0.05		ppm (ug/g)	0.4 max	Pass
Malathion (CAS# 121-75-5)	3/7/2019	< 0.05		ppm (ug/g)	0.2 max	Pass
Metalaxyl (CAS# 57837-19-1)	3/7/2019	< 0.05		ppm (ug/g)	0.2 max	Pass
Methiocarb (CAS# 2032-65-7)	3/7/2019	< 0.05		ppm (ug/g)	0.2 max	Pass
Methomyl (CAS# 16752-77-5)	3/7/2019	< 0.05		ppm (ug/g)	0.4 max	Pass
MGK-264 (CAS#113-48-4)	3/7/2019	< 0.05		ppm (ug/g)	0.2 max	Pass
Myclobutanil (CAS# 88671-89-0)	3/7/2019	< 0.05		ppm (ug/g)	0.2 max	Pass

This report shall not be reproduced without the expressed written consent of ANOVA Laboratory, LLC.

These results pertain only to the representative sample submitted for analysis.

Anova Laboratory, LLC - OLCC Lic. No. 10010169AD1 ORELAP Accreditation No. 4098

Client-Based Solutions and High-Quality, Rapid Results



Anova Laboratory, LLC
 1685 Tucker Road
 Hood River, OR 97031
 541.387.4020 (office)
 866.293.1337 (fax)

Preliminary Analytical Results

Sofresh Farms
 27135 S. Gribble Rd.
 Canby, OR 97013
 Michelle Haworth
 Lic Num: 020-100010499D8

Strain: Betty Hayes
 Matrix: Flower
 CTS Batch #: 1A4010300001E17000005929
 Lot Number: Betty Hayes 8/3/18
 Lot Size (lb): 7.65
 Client CTS Sample ID: SOF_6387
 Collected: 3/1/2019 13:11
 Received: 3/1/2019 16:32
 Anova Sample ID: 079190600201

Analysis Performed	Date Analyzed	Result	fn	Reporting Units	Reporting Criterion	Pass/Fail
Naled (CAS# 300-76-5)	3/7/2019	< 0.20		ppm (ug/g)	0.5 max	Pass
Oxamyl (CAS# 23135-22-0)	3/7/2019	< 0.05		ppm (ug/g)	1 max	Pass
Paclobutrazol (CAS# 76738-62-0)	3/7/2019	< 0.05		ppm (ug/g)	0.4 max	Pass
Permethrins (CAS# 52645-53-1)	3/7/2019	< 0.05		ppm (ug/g)	0.2 max	Pass
Phosmet (CAS#732-11-6)	3/7/2019	< 0.05		ppm (ug/g)	0.2 max	Pass
Piperonyl butoxide (CAS# 51-03-6)	3/7/2019	< 0.05		ppm (ug/g)	2 max	Pass
Prallethrin (CAS# 23031-36-9)	3/7/2019	< 0.05		ppm (ug/g)	0.2 max	Pass
Propiconazole (CAS# 60207-90-1)	3/7/2019	< 0.05		ppm (ug/g)	0.4 max	Pass
Propoxur (CAS# 114-26-1)	3/7/2019	< 0.05		ppm (ug/g)	0.2 max	Pass
Pyrethrins (CAS# 8003-34-7)	3/7/2019	< 0.10		ppm (ug/g)	1 max	Pass
Pyridaben (CAS# 96489-71-3)	3/7/2019	< 0.05		ppm (ug/g)	0.2 max	Pass
Spinosad (Spinosyn A & D) (CAS# 168316-95-8)	3/7/2019	< 0.05		ppm (ug/g)	0.2 max	Pass
Spiromesifen (CAS# 283594-90-1)	3/7/2019	< 0.10		ppm (ug/g)	0.2 max	Pass
Spirotetramat (CAS# 23313-25-1)	3/7/2019	< 0.05		ppm (ug/g)	0.2 max	Pass
Spiroxamine (CAS# 11834-30-8)	3/7/2019	< 0.05		ppm (ug/g)	0.4 max	Pass
Tebuconazole (CAS# 107534-96-3)	3/7/2019	< 0.05		ppm (ug/g)	0.4 max	Pass
Thiacloprid (CAS# 11988-49-9)	3/7/2019	< 0.05		ppm (ug/g)	0.2 max	Pass
Thiamethoxam (CAS# 153719-23-4)	3/7/2019	< 0.05		ppm (ug/g)	0.2 max	Pass
Trifloxystrobin (CAS# 141517-21-7)	3/7/2019	< 0.05		ppm (ug/g)	0.2 max	Pass
AOAC 2007.01 and EN 15662: GC-MS/MS						
Bifenthrin (CAS# 82657-04-3)	3/8/2019	< 0.05		ppm (ug/g)	0.2 max	Pass
Chlorfenapyr (CAS# 122453-73-0)	3/8/2019	< 0.20		ppm (ug/g)	1 max	Pass
Fipronil (CAS# 120068-37-3)	3/8/2019	< 0.05		ppm (ug/g)	0.4 max	Pass
Methyl parathion (CAS# 298-00-0)	3/8/2019	< 0.10		ppm (ug/g)	0.2 max	Pass
Quality Control Analyses						

This report shall not be reproduced without the expressed written consent of ANOVA Laboratory, LLC.

These results pertain only to the representative sample submitted for analysis.

Anova Laboratory, LLC - OLCC Lic. No. 10010169AD1 ORELAP Accreditation No. 4098

Client-Based Solutions and High-Quality, Rapid Results



Anova Laboratory, LLC
 1685 Tucker Road
 Hood River, OR 97031
 541.387.4020 (office)
 866.293.1337 (fax)

Preliminary Analytical Results

Sofresh Farms
 27135 S. Gribble Rd.
 Canby, OR 97013
 Michelle Haworth
 Lic Num: 020-100010499D8

Strain: Betty Hayes
 Matrix: Flower
 CTS Batch #: 1A4010300001E17000005929
 Lot Number: Betty Hayes 8/3/18
 Lot Size (lb): 7.65
 Client CTS Sample ID: SOF_6387
 Collected: 3/1/2019 13:11
 Received: 3/1/2019 16:32
 Anova Sample ID: 079190600201

Analysis Performed	Date Analyzed	Result	fn	Reporting Units	Reporting Criterion	Pass/Fail
LCS_Pesticides (LC-MS/MS)	3/7/2019	positive control		ppm (ug/g)	positive control	Pass
Method Blank_Pesticides (LC-MS/MS)	3/7/2019	clear		ppm (ug/g)	clear	Pass
LCS_Pesticides (GC-MS/MS)	3/8/2019	positive control		ppm (ug/g)	positive control	Pass
Method Blank_Pesticides (GC-MS/MS)	3/8/2019	clear		ppm (ug/g)	clear	Pass

Footnotes:

Acceptance criteria are found in the Oregon Administrative Rule. Analyses are performed in accordance with the most current methods.

Potency results reported on a dry weight basis.

Laboratory Control Sample (LCS) - The percent recovery for target analytes in the batch LCS Quality Control samples support data acceptance.

Method Blank (MB) - The chromatographic regions evaluated for detection were absent of the respective target analytes at or above the Limit of Quantitation (LOQ).

Reviewed By: . , _____

Date: _____

*This report shall not be reproduced without the expressed written consent of ANOVA Laboratory, LLC.
 These results pertain only to the representative sample submitted for analysis.
 Anova Laboratory, LLC - OLCC Lic. No. 10010169AD1 ORELAP Accreditation No. 4098*

Client-Based Solutions and High-Quality, Rapid Results